



ELISA Technologies, Inc.

elisa-tek.com



PAGE 1 OF 1 - REPORT 2109006

REPORT DATE: September 7, 2021

REPORT TO: *Dominic Reale
Goosefoot Acres Inc.
5879 Center Rd.
Valley City, OH 44280
Tel: 440-476-3802
Email: reale.dandyblend@gmail.com*

SUBJECT: *Sample for Gluten Analysis*

RECEIVED: *September 1, 2021; FedEx Ground TRK# 2830 4146 0362*

ANALYTICAL RESULTS: RESULTS APPLY ONLY TO THE SAMPLES AS RECEIVED.

CONTROL #	SAMPLE DESCRIPTION	ANALYSIS IDENTIFICATION	Gluten
2109006-1	<i>Dandy Blend 2# Bag 06-28-2024 Lot # 1261A 04:50</i>	<i>500821 Quantitative Gluten ELISA</i>	Not Detected

NOTE: The ALLER-TEK® Gluten assay is a sandwich ELISA for the quantitative analysis of gluten in raw materials and finished food products. Results are expressed in ppm gluten with a quantifiable range of 2.5-80 ppm; it detects the gliadin and glutenin fractions of gluten. A result of "not detected" indicates that gluten was not detected in the sample (<2.5 ppm). Recovery may be reduced in samples that have been hydrolyzed.

The ALLER-TEK® Gluten ELISA has been validated and certified with license no. 081202 as a Performance Tested Method by the AOAC Research Institute.

Information in italics has been supplied by the customer.

Respectfully Submitted for ELISA Technologies, Inc.

By 
Analyst

By 
Justin Bickford, Scientific Director

End of Report

THIS REPORT SHALL NOT BE REPRODUCED EXCEPT IN ITS ENTIRETY AND WITH THE WRITTEN PERMISSION OF ELISA TECHNOLOGIES, INC.

2501 NW 66th Court, Gainesville, Florida 32653, USA
+1 352 337 3929 info@elisa-tek.com